
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 004476
9:16 04/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:26
PRESSURIZING 3:26
STERILIZING 4:24
136.0 C 220kPa 4:24
136.2 C 222kPa 4:54
136.3 C 225kPa 5:24
136.3 C 224kPa 5:54
136.4 C 222kPa 6:24
136.4 C 224kPa 6:54
136.2 C 220kPa 7:24
136.4 C 224kPa 7:54
136.4 C 224kPa 7:54
VENTING 7:54
AIR DRYING 8:47
CYCLE COMPLETE 8:51

Digital Signature #
303CCBA278C2FD8D
