
Midmark M11 - v1.0.0
Total Cycles: 3742

Name: RITTER 1
Sterilizer ID: V2178198
09/08/2023 12:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.8 F 0.0
2:00 173.4 F 0.8
4:00 192.9 F 4.4
6:00 211.6 F 8.5
8:00 225.7 F 12.2
10:00 231.5 F 9.1
12:00 237.4 F 10.1
14:00 243.2 F 12.4
16:00 241.7 F 11.2
18:00 254.7 F 17.8
20:00 265.9 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.2
1:00 272.5 F 29.0
2:00 271.5 F 28.2
3:00 272.6 F 29.0
4:00 271.7 F 28.3

Min 271.3 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.4 F 0.0
5:00 218.3 F 0.0
10:00 235.9 F 0.0
15:00 225.3 F 0.0
20:00 222.2 F 0.0
25:00 228.4 F 0.0
30:00 230.4 F 0.0
Drying Complete

FILLING: 1:54
HEATING: 21:34
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:00:56
CYCLE COMPLETE

09/08/2023 01:21 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____