

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1448

Name: RITTER2  
Sterilizer ID: V2313228  
09/02/2022 11:29 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 126.6 F 0.0  
2:00 128.9 F 0.0  
4:00 143.6 F 1.2  
6:00 175.1 F 4.3  
8:00 205.2 F 7.6  
10:00 221.1 F 12.3  
12:00 226.3 F 7.5  
14:00 238.0 F 12.5  
16:00 236.8 F 10.0  
18:00 238.1 F 8.9  
20:00 248.9 F 15.0  
22:00 260.5 F 21.5  
24:00 270.9 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.8  
1:00 272.5 F 29.6  
2:00 271.8 F 28.9  
3:00 271.4 F 29.2  
4:00 272.6 F 29.5

Min 271.2 F 28.6  
Max 272.6 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.8 F 0.0  
5:00 211.5 F 0.0  
10:00 220.6 F 0.0  
15:00 220.5 F 0.0  
20:00 217.1 F 0.0  
25:00 215.4 F 0.0  
30:00 215.1 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 24:47  
STERILIZING: 4:00  
VENTING: 3:02  
DRYING: 30:00  
TOTAL CYCLE: 01:02:46  
CYCLE COMPLETE

09/02/2022 12:31 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_