

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.6uS / 0.3ppm  
CYCLE NUMBER 001434  
15:22 14/08/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:02  
PRESSURIZING 3:07  
STERILIZING 4:13  
136.4 C 224kPa 4:13  
136.3 C 225kPa 4:43  
136.4 C 226kPa 5:13  
136.4 C 225kPa 5:43  
136.1 C 222kPa 6:13  
136.4 C 226kPa 6:43  
136.4 C 226kPa 7:13  
136.4 C 225kPa 7:43  
136.4 C 225kPa 7:43  
VENTING 7:44  
AIR DRYING 8:40  
CYCLE COMPLETE 9:25  
-----

Digital Signature #  
FA89F18BFC16AB7A  
-----