
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076
Unit # : 000
WATER QUALITY
2.1uS / 1.3ppm
CYCLE NUMBER 000022
5:44 14/01/2020
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 6:18
PRESSURIZING 8:27
STERILIZING 10:38
136.5 C 228kPa 10:38
136.4 C 227kPa 11:09
136.2 C 226kPa 11:39
136.4 C 225kPa 12:09
136.3 C 226kPa 12:39
136.4 C 227kPa 13:09
136.4 C 227kPa 13:39
136.4 C 227kPa 14:09
VENTING 14:09
AIR DRYING 15:25
CYCLE COMPLETE 45:25

Digital Signature #
3F359FDA752DAE15
