

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 000215  
13:43 08/07/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:56  
PRESSURIZING 2:58  
STERILIZING 3:55  
136.4 C 222kPa 3:55  
136.4 C 223kPa 4:25  
136.4 C 221kPa 4:55  
136.3 C 224kPa 5:25  
136.3 C 223kPa 5:55  
136.4 C 219kPa 6:25  
136.2 C 222kPa 6:55  
136.4 C 224kPa 7:25  
136.4 C 224kPa 7:25  
VENTING 7:25  
AIR DRYING 8:14  
CYCLE COMPLETE 8:23

-----  
Digital Signature #  
1BB8E216328FA2FA  
-----