
Midmark M11 - v1.0.0
Total Cycles: 1573

Name: RITTER 1
Sterilizer ID: V2178198
01/31/2022 01:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.4 F 0.0
2:00 173.9 F 3.3
4:00 192.6 F 7.2
6:00 205.4 F 11.8
8:00 219.8 F 5.1
10:00 230.5 F 9.9
12:00 238.7 F 8.6
14:00 243.6 F 12.3
16:00 240.8 F 10.6
18:00 253.8 F 17.1
20:00 265.3 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.8 F 29.1
2:00 271.9 F 28.4
3:00 271.4 F 28.3
4:00 272.6 F 29.1

Min 271.2 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.0 F 0.0
5:00 211.8 F 0.0
10:00 239.3 F 0.0
15:00 230.7 F 0.0
20:00 228.5 F 0.0
25:00 227.8 F 0.0
30:00 225.9 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 21:52
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:12
CYCLE COMPLETE

01/31/2022 02:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____