
Midmark M11 - v1.0.5
Total Cycles: 1551

Name: RITTER2
Sterilizer ID: V2313228
10/03/2022 09:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.3 F 0.0
2:00 189.2 F 0.2
4:00 201.0 F 3.9
6:00 221.7 F 9.4
8:00 233.5 F 8.5
10:00 240.9 F 12.5
12:00 240.8 F 11.2
14:00 240.8 F 10.7
16:00 254.7 F 18.1
18:00 267.0 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.4 F 29.1
2:00 272.3 F 29.5
3:00 271.9 F 28.8
4:00 272.5 F 29.5

Min 271.4 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.4 F 0.0
5:00 230.0 F 0.0
10:00 237.8 F 0.0
15:00 241.3 F 0.0
20:00 228.9 F 0.0
25:00 222.3 F 0.0
30:00 220.5 F 0.0

Drying Complete

FILLING: 1:40
HEATING: 19:23
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:10
CYCLE COMPLETE

10/03/2022 10:37 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____