
Midmark M11 - v1.0.0
Total Cycles: 2823

Name: RITTER 1
Sterilizer ID: V2178198
01/30/2023 09:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.5 F 0.0
2:00 174.9 F 2.5
4:00 181.3 F 6.6
6:00 192.9 F 11.2
8:00 211.7 F 4.3
10:00 227.9 F 9.3
12:00 235.8 F 9.7
14:00 239.4 F 11.1
16:00 237.1 F 8.9
18:00 250.4 F 15.5
20:00 262.0 F 22.1
22:00 272.5 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.8 F 29.2
2:00 271.8 F 28.4
3:00 271.9 F 28.9
4:00 272.6 F 29.1

Min 271.4 F 28.3
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.8 F 0.0
5:00 221.8 F 0.0
10:00 236.6 F 0.0
15:00 229.1 F 0.0
20:00 226.1 F 0.0
25:00 228.9 F 0.0
30:00 229.6 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 22:27
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:00:43
CYCLE COMPLETE

01/30/2023 10:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____