
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 004934
14:35 09/01/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:05
STERILIZING 4:14
136.4 C 221kPa 4:14
136.3 C 225kPa 4:45
136.4 C 225kPa 5:15
136.2 C 218kPa 5:45
136.4 C 223kPa 6:15
136.4 C 222kPa 6:45
136.1 C 220kPa 7:15
136.4 C 223kPa 7:45
136.4 C 223kPa 7:45
VENTING 7:45
AIR DRYING 8:36
CYCLE COMPLETE 21:04

Digital Signature #
BA7A3900921BD505
