
Midmark M11 - v1.0.5
Total Cycles: 3747

Name: RITTER2
Sterilizer ID: V2313228
06/28/2024 02:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.0 F 0.0
2:00 146.1 F 2.4
4:00 169.8 F 6.0
6:00 193.8 F 10.4
8:00 215.5 F 5.4
10:00 222.7 F 8.0
12:00 235.9 F 12.6
14:00 236.4 F 9.8
16:00 237.8 F 9.3
18:00 248.8 F 15.0
20:00 260.6 F 21.6
22:00 271.4 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.6 F 29.3
2:00 271.5 F 29.1
3:00 272.8 F 29.3
4:00 271.3 F 28.8

Min 271.3 F 28.7
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.0 F 0.0
5:00 230.9 F 0.0
10:00 202.5 F 0.0
15:00 182.3 F 0.0
20:00 176.6 F 0.0
25:00 174.1 F 0.0
30:00 172.1 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 22:38
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:01:34
CYCLE COMPLETE

06/28/2024 03:23 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____