
Midmark M11 - v1.0.0
Total Cycles: 607

Name: MIDMARK
Sterilizer ID: V2178198
01/18/2021 07:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.7 F 0.0
2:00 72.0 F 0.0
4:00 90.0 F 0.1
6:00 133.7 F 0.1
8:00 153.5 F 0.2
10:00 190.2 F 0.4
12:00 204.1 F 3.9
14:00 213.4 F 7.0
16:00 223.7 F 10.4
18:00 230.7 F 9.3
20:00 235.2 F 10.0
22:00 236.4 F 9.0
24:00 244.7 F 13.0
26:00 242.5 F 11.6
28:00 255.4 F 18.1
30:00 266.4 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.2 F 28.6
2:00 272.3 F 29.0
3:00 271.8 F 28.3
4:00 272.7 F 29.1

Min 271.4 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.4 F 0.0
5:00 203.0 F 0.0
10:00 212.3 F 0.0
15:00 215.2 F 0.0
20:00 217.3 F 0.0
25:00 221.3 F 0.0
30:00 227.1 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 31:31
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:09:26
CYCLE COMPLETE

01/18/2021 08:47 AM
ACCEPTED: X
REJECTED:
Operator: U01