

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4364

Name: RITTER2  
Sterilizer ID: V2313228  
02/20/2025 01:50 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 148.8 F 0.0  
2:00 151.7 F 2.1  
4:00 159.9 F 5.6  
6:00 189.0 F 10.1  
8:00 213.5 F 6.3  
10:00 225.6 F 10.0  
12:00 229.5 F 8.3  
14:00 240.8 F 12.9  
16:00 240.9 F 11.6  
18:00 253.7 F 18.2  
20:00 265.0 F 25.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 29.5  
1:00 273.1 F 30.3  
2:00 272.6 F 30.2  
3:00 272.3 F 30.1  
4:00 272.2 F 30.2

Min 270.8 F 29.4  
Max 273.1 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.4 F 0.0  
5:00 235.5 F 0.0  
10:00 229.5 F 0.0  
15:00 230.4 F 0.0  
20:00 228.3 F 0.0  
25:00 227.8 F 0.0  
30:00 226.4 F 0.0  
Drying Complete

FILLING: 1:07  
HEATING: 21:51  
STERILIZING: 4:00  
VENTING: 3:40  
DRYING: 30:00  
TOTAL CYCLE: 01:00:38  
CYCLE COMPLETE

02/20/2025 02:50 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_