
Midmark M11 - v1.0.0
Total Cycles: 3975

Name: RITTER 1
Sterilizer ID: V2178198
11/02/2023 02:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.9 F 0.0
2:00 170.5 F 3.7
4:00 192.4 F 8.2
6:00 217.7 F 9.1
8:00 227.6 F 8.1
10:00 239.0 F 12.7
12:00 240.2 F 10.8
14:00 237.4 F 9.0
16:00 252.0 F 16.0
18:00 264.3 F 23.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.3 F 28.2
2:00 272.7 F 28.7
3:00 271.5 F 27.9
4:00 272.7 F 28.6

Min 271.5 F 27.8
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.5 F 0.0
5:00 240.9 F 0.0
10:00 226.6 F 0.0
15:00 221.1 F 0.0
20:00 229.0 F 0.0
25:00 230.3 F 0.0
30:00 226.5 F 0.0

Drying Complete

FILLING: 1:33
HEATING: 19:50
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:58:43
CYCLE COMPLETE

11/02/2023 03:50 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____