
Midmark M11 - v1.0.0
Total Cycles: 4006

Name: RITTER 1
Sterilizer ID: V2178198
11/10/2023 11:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.4 F 0.0
2:00 179.2 F 2.8
4:00 180.1 F 6.4
6:00 196.5 F 11.0
8:00 218.7 F 4.4
10:00 226.9 F 8.9
12:00 237.4 F 12.6
14:00 237.6 F 10.5
16:00 238.5 F 8.4
18:00 248.7 F 14.8
20:00 260.0 F 20.9
22:00 270.2 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.6 F 29.2
2:00 272.3 F 28.8
3:00 271.3 F 28.4
4:00 272.0 F 29.0

Min 271.2 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.4 F 0.0
5:00 231.7 F 0.0
10:00 229.8 F 0.0
15:00 227.5 F 0.0
20:00 224.2 F 0.0
25:00 221.4 F 0.0
30:00 219.1 F 0.0
Drying Complete

FILLING: 1:24
HEATING: 22:57
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:02:04
CYCLE COMPLETE

11/10/2023 12:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____