

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4808

Name: RITTER 1  
Sterilizer ID: V2178198  
08/13/2024 08:37 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.2 F 0.0  
2:00 72.2 F 0.0  
4:00 84.3 F 0.1  
6:00 118.5 F 0.2  
8:00 148.1 F 1.5  
10:00 176.2 F 4.1  
12:00 202.7 F 7.7  
14:00 216.3 F 11.4  
16:00 221.6 F 4.4  
18:00 230.4 F 7.9  
20:00 238.0 F 12.1  
22:00 234.6 F 8.3  
24:00 245.1 F 12.1  
26:00 243.8 F 11.9  
28:00 255.5 F 18.0  
30:00 266.2 F 24.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.1  
1:00 272.6 F 28.8  
2:00 271.9 F 28.1  
3:00 271.8 F 28.5  
4:00 272.7 F 28.8

Min 271.3 F 27.9  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.0 F 0.0  
5:00 210.6 F 0.0  
10:00 204.2 F 0.0  
15:00 230.9 F 0.0  
20:00 225.2 F 0.0  
25:00 224.5 F 0.0  
30:00 211.4 F 0.0  
Drying Complete

FILLING: 0:42  
HEATING: 31:45  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 01:09:47  
CYCLE COMPLETE

08/13/2024 09:46 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_