
Midmark M11 - v1.0.0
Total Cycles: 3850

Name: RITTER 1
Sterilizer ID: V2178198
10/02/2023 01:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 213.3 F 0.0
2:00 201.8 F 4.3
4:00 211.1 F 10.2
6:00 224.7 F 5.6
8:00 238.4 F 11.2
10:00 236.3 F 8.8
12:00 239.3 F 9.5
14:00 249.8 F 14.8
16:00 262.5 F 22.0
18:00 271.8 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.9
1:00 271.7 F 27.8
2:00 271.9 F 28.1
3:00 272.3 F 28.3
4:00 272.7 F 28.7

Min 271.5 F 27.8
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.5 F 0.0
5:00 216.1 F 0.0
10:00 231.8 F 0.0
15:00 231.5 F 0.0
20:00 225.4 F 0.0
25:00 229.2 F 0.0
30:00 222.0 F 0.0
Drying Complete

FILLING: 2:19
HEATING: 18:06
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:57:41
CYCLE COMPLETE

10/02/2023 02:49 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____