

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4123

Name: RITTER2  
Sterilizer ID: V2313228  
11/14/2024 07:11 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 100.0 F 0.0  
2:00 106.0 F 0.0  
4:00 131.6 F 0.6  
6:00 168.3 F 4.2  
8:00 192.3 F 8.0  
10:00 209.8 F 12.4  
12:00 218.7 F 6.0  
14:00 230.0 F 10.5  
16:00 232.3 F 8.0  
18:00 243.3 F 13.1  
20:00 242.9 F 12.0  
22:00 255.2 F 18.4  
24:00 266.2 F 25.2

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 28.9  
1:00 272.7 F 29.4  
2:00 272.6 F 29.6  
3:00 271.8 F 28.9  
4:00 272.4 F 29.1

Min 271.7 F 28.7  
Max 272.8 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 176.0 F 0.0  
5:00 210.4 F 0.0  
10:00 202.1 F 0.0  
15:00 198.2 F 0.0  
20:00 194.0 F 0.0  
25:00 190.1 F 0.0  
30:00 189.6 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 25:28  
STERILIZING: 4:00  
VENTING: 3:32  
DRYING: 30:00  
TOTAL CYCLE: 01:04:03  
CYCLE COMPLETE

11/14/2024 08:15 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_