

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 000255  
13:38 21/07/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:43  
PRESSURIZING 3:44  
STERILIZING 4:59  
136.4 C 221kPa 4:59  
136.3 C 221kPa 5:29  
136.0 C 222kPa 5:59  
136.1 C 222kPa 6:29  
136.2 C 223kPa 6:59  
136.3 C 223kPa 7:29  
136.3 C 224kPa 7:59  
136.3 C 222kPa 8:29  
136.3 C 222kPa 8:29  
VENTING 8:29  
AIR DRYING 9:23  
CYCLE COMPLETE 9:28

-----  
Digital Signature #  
78AE4BEFC92BA121  
-----