
Midmark M11 - v1.0.5
Total Cycles: 4069

Name: RITTER2
Sterilizer ID: V2313228
10/30/2024 04:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 109.8 F 0.0
2:00 116.1 F 0.0
4:00 146.2 F 0.1
6:00 170.9 F 3.2
8:00 196.1 F 7.2
10:00 209.4 F 11.7
12:00 218.3 F 5.7
14:00 230.7 F 10.3
16:00 233.4 F 8.4
18:00 243.3 F 13.1
20:00 242.8 F 12.0
22:00 255.4 F 18.5
24:00 266.5 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.7 F 29.4
2:00 271.6 F 28.8
3:00 272.7 F 29.4
4:00 271.7 F 28.6

Min 271.5 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.9 F 0.0
5:00 196.4 F 0.0
10:00 200.5 F 0.0
15:00 200.4 F 0.0
20:00 198.2 F 0.0
25:00 194.3 F 0.0
30:00 190.8 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 25:34
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:03:56
CYCLE COMPLETE

10/30/2024 05:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____