
Midmark M11 - v1.0.5
Total Cycles: 2797

Name: RITTER2
Sterilizer ID: V2313228
08/25/2023 11:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.2 F 0.0
2:00 168.3 F 0.1
4:00 197.9 F 3.0
6:00 220.9 F 7.1
8:00 220.6 F 11.6
10:00 227.1 F 6.9
12:00 239.7 F 11.9
14:00 236.8 F 9.7
16:00 238.5 F 9.6
18:00 248.7 F 14.8
20:00 260.0 F 21.2
22:00 270.1 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.8 F 29.6
2:00 271.9 F 28.9
3:00 271.3 F 29.1
4:00 272.7 F 29.5

Min 271.0 F 28.6
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.7 F 0.0
5:00 244.5 F 0.0
10:00 234.4 F 0.0
15:00 219.4 F 0.0
20:00 212.0 F 0.0
25:00 212.3 F 0.0
30:00 216.4 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 22:57
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:02:01
CYCLE COMPLETE

08/25/2023 12:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____