
Midmark M11 - v1.0.5
Total Cycles: 880

Name: RITTER2
Sterilizer ID: V2313228
03/14/2022 02:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.2 F 0.0
2:00 184.4 F 4.0
4:00 203.9 F 9.5
6:00 221.2 F 7.1
8:00 232.0 F 11.8
10:00 235.7 F 9.9
12:00 236.4 F 8.8
14:00 250.9 F 16.3
16:00 263.2 F 23.5
18:00 271.4 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.3 F 29.7
2:00 271.2 F 29.1
3:00 272.7 F 29.9
4:00 271.8 F 29.3

Min 271.2 F 29.0
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.9 F 0.0
5:00 231.7 F 0.0
10:00 221.5 F 0.0
15:00 220.1 F 0.0
20:00 226.5 F 0.0
25:00 231.0 F 0.0
30:00 213.9 F 0.0
Drying Complete

FILLING: 2:13
HEATING: 18:04
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:57:26
CYCLE COMPLETE

03/14/2022 03:30 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____