

-----  
Midmark M11 - v1.0.0  
Total Cycles: 682

Name: MIDMARK  
Sterilizer ID: V2178198  
02/19/2021 01:46 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 106.2 F 0.0  
2:00 123.8 F 0.1  
4:00 138.8 F 0.2  
6:00 173.1 F 3.1  
8:00 197.6 F 7.2  
10:00 215.4 F 12.2  
12:00 223.0 F 6.9  
14:00 236.1 F 12.3  
16:00 236.6 F 9.5  
18:00 236.7 F 8.5  
20:00 249.0 F 14.7  
22:00 261.8 F 21.9  
24:00 272.4 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 271.5 F 28.3  
2:00 272.1 F 28.6  
3:00 272.5 F 29.0  
4:00 272.0 F 29.0

Min 271.4 F 28.3  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.5 F 0.0  
5:00 236.7 F 0.0  
10:00 223.4 F 0.0  
15:00 222.2 F 0.0  
20:00 227.7 F 0.0  
25:00 229.6 F 0.0  
30:00 222.9 F 0.0  
Drying Complete

FILLING: 1:05  
HEATING: 24:11  
STERILIZING: 4:00  
VENTING: 2:56  
DRYING: 30:00  
TOTAL CYCLE: 01:02:13  
CYCLE COMPLETE

02/19/2021 02:48 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_