

-----  
Midmark M11 - v1.0.5  
Total Cycles: 13

Name: RITTER2  
Sterilizer ID: V2313228  
04/22/2021 03:39 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 175.6 F 0.0  
2:00 180.0 F 4.0  
4:00 203.9 F 9.2  
6:00 222.1 F 7.6  
8:00 231.3 F 10.9  
10:00 234.8 F 9.3  
12:00 238.3 F 10.0  
14:00 249.9 F 15.8  
16:00 262.3 F 22.8  
18:00 272.6 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.9  
1:00 272.4 F 29.6  
2:00 271.5 F 28.8  
3:00 272.1 F 29.8  
4:00 272.0 F 29.3  
  
Min 271.2 F 28.7  
Max 272.6 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.1 F 0.0  
5:00 239.5 F 0.0  
10:00 218.5 F 0.0  
15:00 201.7 F 0.0  
20:00 191.2 F 0.0  
25:00 187.7 F 0.0  
30:00 189.8 F 0.0  
Drying Complete

FILLING: 2:34  
HEATING: 18:15  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:58:01  
CYCLE COMPLETE

04/26/2021 06:06 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02