
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001468
14:00 06/04/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:37
PRESSURIZING 3:38
STERILIZING 4:36
136.4 C 220kPa 4:36
136.3 C 223kPa 5:06
136.2 C 221kPa 5:36
136.3 C 223kPa 6:06
136.3 C 224kPa 6:36
136.3 C 223kPa 7:06
136.4 C 224kPa 7:36
136.3 C 223kPa 8:06
136.3 C 223kPa 8:06
VENTING 8:06
AIR DRYING 9:07
CYCLE COMPLETE 9:09

Digital Signature #
1B37DA0C4FE2A7EC
