

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1925

Name: RITTER 1  
Sterilizer ID: V2178198  
05/10/2022 11:53 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 180.5 F 0.0  
2:00 177.4 F 3.6  
4:00 183.4 F 7.7  
6:00 203.1 F 12.4  
8:00 220.7 F 6.9  
10:00 235.9 F 11.9  
12:00 233.2 F 8.3  
14:00 242.9 F 11.3  
16:00 244.3 F 12.4  
18:00 256.6 F 18.9  
20:00 267.5 F 25.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.6 F 29.0  
2:00 271.9 F 28.4  
3:00 271.6 F 28.5  
4:00 272.6 F 29.0  
  
Min 271.4 F 28.1  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.0 F 0.0  
5:00 228.2 F 0.0  
10:00 208.3 F 0.0  
15:00 227.3 F 0.0  
20:00 235.5 F 0.0  
25:00 228.3 F 0.0  
30:00 227.6 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 21:27  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:59:37  
CYCLE COMPLETE

05/10/2022 12:52 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_