
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 004442
15:49 26/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:33
PRESSURIZING 2:34
STERILIZING 3:34
136.0 C 220kPa 3:34
136.4 C 225kPa 4:04
136.4 C 224kPa 4:34
136.1 C 224kPa 5:04
136.1 C 222kPa 5:34
136.2 C 223kPa 6:04
136.2 C 223kPa 6:34
136.3 C 225kPa 7:04
136.3 C 225kPa 7:04
VENTING 7:04
AIR DRYING 8:16
CYCLE COMPLETE 13:44

Digital Signature #
EF952AD388F49D1D
