
Midmark M11 - v1.0.5
Total Cycles: 2722

Name: RITTER2
Sterilizer ID: V2313228
08/01/2023 04:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 211.3 F 0.0
2:00 204.7 F 4.1
4:00 209.0 F 10.0
6:00 223.1 F 6.5
8:00 236.3 F 11.9
10:00 237.1 F 10.2
12:00 236.9 F 9.2
14:00 251.5 F 16.5
16:00 263.8 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.6 F 29.4
2:00 271.6 F 29.1
3:00 272.5 F 29.4
4:00 271.5 F 28.9

Min 271.5 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.7 F 0.0
5:00 228.3 F 0.0
10:00 237.3 F 0.0
15:00 235.1 F 0.0
20:00 225.1 F 0.0
25:00 225.7 F 0.0
30:00 226.0 F 0.0

Drying Complete

FILLING: 3:20
HEATING: 17:59
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:34
CYCLE COMPLETE

08/01/2023 05:10 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____