

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4106

Name: RITTER 1  
Sterilizer ID: V2178198  
12/11/2023 04:19 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 156.4 F 0.0  
2:00 159.5 F 2.7  
4:00 182.9 F 6.4  
6:00 205.2 F 11.5  
8:00 221.7 F 5.3  
10:00 230.7 F 10.2  
12:00 233.3 F 7.9  
14:00 243.5 F 12.8  
16:00 242.8 F 11.6  
18:00 255.6 F 18.1  
20:00 266.7 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.2  
1:00 272.7 F 29.2  
2:00 271.7 F 28.9  
3:00 272.1 F 28.7  
4:00 272.7 F 29.2  
  
Min 271.0 F 28.1  
Max 272.8 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 176.4 F 0.0  
5:00 257.5 F 0.0  
10:00 237.8 F 0.0  
15:00 227.2 F 0.0  
20:00 230.7 F 0.0  
25:00 227.6 F 0.0  
30:00 223.4 F 0.0  
Drying Complete

FILLING: 1:41  
HEATING: 21:33  
STERILIZING: 4:00  
VENTING: 3:41  
DRYING: 30:00  
TOTAL CYCLE: 01:00:55  
CYCLE COMPLETE

12/11/2023 05:19 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_