

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 000143  
16:13 16/06/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:54  
PRESSURIZING 2:56  
STERILIZING 3:56  
136.4 C 220kPa 3:56  
136.4 C 225kPa 4:26  
136.3 C 222kPa 4:56  
136.4 C 222kPa 5:26  
136.4 C 225kPa 5:56  
136.4 C 220kPa 6:26  
136.1 C 220kPa 6:56  
136.3 C 223kPa 7:26  
136.3 C 223kPa 7:26  
VENTING 7:26  
AIR DRYING 8:16  
CYCLE COMPLETE 8:39

-----  
Digital Signature #  
E868FF47C37C0A39  
-----