
Midmark M11 - v1.0.0
Total Cycles: 4363

Name: RITTER 1
Sterilizer ID: V2178198
03/07/2024 10:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.5 F 0.0
2:00 156.1 F 1.8
4:00 158.4 F 4.5
6:00 171.8 F 8.0
8:00 192.9 F 12.5
10:00 212.8 F 6.3
12:00 221.9 F 10.8
14:00 230.2 F 6.7
16:00 237.3 F 11.9
18:00 237.4 F 9.5
20:00 250.2 F 15.5
22:00 260.9 F 21.6
24:00 270.7 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.1
1:00 272.4 F 29.2
2:00 272.7 F 29.6
3:00 271.8 F 29.4
4:00 272.3 F 29.2

Min 271.7 F 29.0
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.4 F 0.0
5:00 226.1 F 0.0
10:00 228.3 F 0.0
15:00 226.1 F 0.0
20:00 228.5 F 0.0
25:00 230.5 F 0.0
30:00 229.7 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 25:32
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:04:16
CYCLE COMPLETE

03/07/2024 11:22 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____