
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 007247
9:27 23/04/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:22
PRESSURIZING 2:25
STERILIZING 3:27
136.0 C 220kPa 3:27
136.3 C 223kPa 3:57
136.4 C 224kPa 4:27
136.3 C 224kPa 4:57
136.2 C 221kPa 5:27
136.4 C 222kPa 5:57
136.4 C 222kPa 6:27
136.4 C 224kPa 6:57
136.4 C 224kPa 6:57
VENTING 6:57
AIR DRYING 7:50
CYCLE COMPLETE 37:16

Digital Signature #
481F555B0C03DFFA
