
Midmark M11 - v1.0.5
Total Cycles: 3866

Name: RITTER2
Sterilizer ID: V2313228
08/16/2024 09:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.0 F 0.0
2:00 74.6 F 0.0
4:00 84.6 F 0.0
6:00 108.0 F 0.7
8:00 138.9 F 3.1
10:00 167.7 F 5.8
12:00 187.8 F 9.4
14:00 198.6 F 13.0
16:00 211.9 F 4.9
18:00 222.3 F 8.3
20:00 231.5 F 12.1
22:00 231.7 F 8.3
24:00 242.6 F 13.2
26:00 241.2 F 11.4
28:00 252.2 F 17.0
30:00 262.2 F 22.8
32:00 271.6 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.7 F 29.8
2:00 271.9 F 29.1
3:00 271.5 F 29.3
4:00 272.4 F 29.7

Min 271.2 F 28.8
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.4 F 0.0
5:00 214.1 F 0.0
10:00 190.1 F 0.0
15:00 185.0 F 0.0
20:00 182.7 F 0.0
25:00 182.5 F 0.0
30:00 173.4 F 0.0

Drying Complete

FILLING: 0:59
HEATING: 32:39
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:11:04
CYCLE COMPLETE

08/16/2024 10:34 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____