

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4817

Name: RITTER 1  
Sterilizer ID: V2178198  
08/14/2024 02:15 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 140.0 F 0.0  
2:00 145.6 F 0.7  
4:00 168.4 F 1.4  
6:00 190.4 F 4.7  
8:00 207.3 F 7.9  
10:00 217.2 F 12.0  
12:00 221.7 F 5.5  
14:00 232.2 F 9.8  
16:00 236.6 F 9.8  
18:00 242.6 F 12.0  
20:00 240.5 F 10.6  
22:00 253.1 F 16.8  
24:00 263.6 F 22.9  
26:00 272.7 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.2  
1:00 272.2 F 28.8  
2:00 272.5 F 28.8  
3:00 272.1 F 28.5  
4:00 271.6 F 28.1

Min 271.4 F 28.1  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.6 F 0.0  
5:00 220.5 F 0.0  
10:00 215.2 F 0.0  
15:00 216.0 F 0.0  
20:00 208.4 F 0.0  
25:00 211.8 F 0.0  
30:00 213.5 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 26:26  
STERILIZING: 4:00  
VENTING: 3:52  
DRYING: 30:00  
TOTAL CYCLE: 01:05:15  
CYCLE COMPLETE

08/14/2024 03:20 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_