
Midmark M11 - v1.0.0
Total Cycles: 4977

Name: RITTER 1
Sterilizer ID: V2178198
10/08/2024 01:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 89.5 F 0.0
2:00 90.1 F 0.0
4:00 107.9 F 0.0
6:00 128.2 F 0.3
8:00 158.0 F 1.8
10:00 198.5 F 4.8
12:00 211.1 F 8.0
14:00 221.9 F 11.5
16:00 224.3 F 5.1
18:00 229.9 F 9.8
20:00 237.7 F 9.8
22:00 240.5 F 11.4
24:00 237.2 F 9.3
26:00 249.7 F 15.3
28:00 260.3 F 21.2
30:00 270.0 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.5 F 29.2
2:00 271.9 F 28.6
3:00 271.5 F 28.8
4:00 272.4 F 29.0

Min 271.2 F 28.4
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.3 F 0.0
5:00 209.0 F 0.0
10:00 218.2 F 0.0
15:00 231.4 F 0.0
20:00 229.6 F 0.0
25:00 221.5 F 0.0
30:00 220.6 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 30:59
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:09:36
CYCLE COMPLETE

10/08/2024 03:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____