

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001692  
11:26 05/02/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:44  
PRESSURIZING 2:50  
STERILIZING 4:04  
136.4 C 225kPa 4:04  
136.3 C 225kPa 4:34  
136.5 C 224kPa 5:04  
136.1 C 223kPa 5:34  
136.4 C 225kPa 6:04  
136.4 C 222kPa 6:34  
136.3 C 224kPa 7:04  
136.4 C 225kPa 7:34  
136.4 C 225kPa 7:34  
VENTING 7:34  
AIR DRYING 8:24  
CYCLE COMPLETE 8:28  
-----

Digital Signature #  
F5DC305E4D1B4821  
-----