
Midmark M11 - v1.0.5
Total Cycles: 2842

Name: RITTER2
Sterilizer ID: V2313228
09/07/2023 09:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.7 F 0.0
2:00 140.9 F 0.0
4:00 142.2 F 0.3
6:00 168.3 F 3.4
8:00 191.0 F 7.0
10:00 212.2 F 11.0
12:00 219.4 F 6.1
14:00 230.3 F 10.0
16:00 232.5 F 9.2
18:00 240.8 F 12.3
20:00 240.1 F 11.0
22:00 251.9 F 17.0
24:00 262.6 F 23.0
26:00 272.1 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 272.6 F 29.9
2:00 272.7 F 30.0
3:00 272.0 F 29.9
4:00 271.5 F 29.5

Min 271.0 F 29.0
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.6 F 0.0
5:00 208.3 F 0.0
10:00 223.1 F 0.0
15:00 223.2 F 0.0
20:00 218.8 F 0.0
25:00 217.2 F 0.0
30:00 213.0 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 26:23
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:04:47
CYCLE COMPLETE

09/07/2023 10:51 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____