
Midmark M11 - v1.0.5
Total Cycles: 528

Name: RITTER2
Sterilizer ID: V2313228
11/09/2021 12:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 137.2 F 0.0
2:00 142.2 F 2.3
4:00 166.9 F 6.1
6:00 193.1 F 11.0
8:00 213.5 F 5.0
10:00 226.1 F 10.3
12:00 230.3 F 7.7
14:00 242.1 F 13.3
16:00 242.5 F 12.2
18:00 255.2 F 19.0
20:00 266.9 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.3
1:00 272.1 F 30.1
2:00 272.1 F 29.5
3:00 271.6 F 29.7
4:00 272.5 F 30.0

Min 271.3 F 29.2
Max 272.9 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.8 F 0.0
5:00 215.1 F 0.0
10:00 231.5 F 0.0
15:00 223.9 F 0.0
20:00 221.8 F 0.0
25:00 221.4 F 0.0
30:00 220.0 F 0.0
Drying Complete

FILLING: 1:29
HEATING: 21:21
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 00:59:52
CYCLE COMPLETE

11/09/2021 01:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____