
Midmark M11 - v1.0.0
Total Cycles: 4825

Name: RITTER 1
Sterilizer ID: V2178198
08/16/2024 10:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.6 F 0.0
2:00 147.9 F 0.5
4:00 163.8 F 1.0
6:00 184.8 F 2.6
8:00 202.1 F 6.5
10:00 218.7 F 10.5
12:00 229.0 F 6.9
14:00 232.1 F 9.4
16:00 239.7 F 10.7
18:00 242.6 F 11.6
20:00 238.9 F 9.7
22:00 251.5 F 15.8
24:00 262.2 F 21.9
26:00 271.5 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.5 F 28.9
2:00 272.1 F 28.4
3:00 271.3 F 28.2
4:00 272.2 F 28.9

Min 271.3 F 28.1
Max 272.5 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.0 F 0.0
5:00 225.3 F 0.0
10:00 210.6 F 0.0
15:00 209.1 F 0.0
20:00 224.8 F 0.0
25:00 231.5 F 0.0
30:00 225.4 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 26:42
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:05:18
CYCLE COMPLETE

08/16/2024 11:21 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____