
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000394
11:49 25/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:08
PRESSURIZING 3:12
STERILIZING 4:14
136.0 C 228kPa 4:14
136.4 C 233kPa 4:44
135.9 C 229kPa 5:14
136.3 C 231kPa 5:44
136.3 C 233kPa 6:14
136.4 C 233kPa 6:44
136.4 C 233kPa 7:14
136.2 C 233kPa 7:44
136.2 C 233kPa 7:44
VENTING 7:44
AIR DRYING 8:36
CYCLE COMPLETE 8:38

Digital Signature #
DFDA86112D45FC5C
