

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2223

Name: RITTER 1  
Sterilizer ID: V2178198  
08/15/2022 01:06 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 167.9 F 0.0  
2:00 164.6 F 3.8  
4:00 182.1 F 7.9  
6:00 210.3 F 12.3  
8:00 223.4 F 6.2  
10:00 233.7 F 11.1  
12:00 231.5 F 7.2  
14:00 243.5 F 13.0  
16:00 240.9 F 11.1  
18:00 253.4 F 17.3  
20:00 264.4 F 23.8  
22:00 271.4 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 272.7 F 29.2  
2:00 271.6 F 28.5  
3:00 272.6 F 29.4  
4:00 272.5 F 29.0

Min 271.5 F 28.5  
Max 272.8 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.8 F 0.0  
5:00 228.9 F 0.0  
10:00 229.9 F 0.0  
15:00 234.7 F 0.0  
20:00 229.6 F 0.0  
25:00 229.2 F 0.0  
30:00 224.3 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 22:01  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:00:11  
CYCLE COMPLETE

08/15/2022 02:06 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_