

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002494  
7:46 22/10/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:16  
PRESSURIZING 2:24  
STERILIZING 3:18  
136.0 C 220kPa 3:18  
136.1 C 220kPa 3:48  
136.4 C 223kPa 4:18  
136.4 C 223kPa 4:48  
136.3 C 222kPa 5:18  
136.4 C 222kPa 5:48  
136.4 C 223kPa 6:18  
136.4 C 222kPa 6:48  
136.4 C 222kPa 6:48  
VENTING 6:48  
AIR DRYING 7:38  
CYCLE COMPLETE 7:50

-----  
Digital Signature #  
75BC76846E5E31DB  
-----