
Midmark M11 - v1.0.0
Total Cycles: 3696

Name: RITTER 1
Sterilizer ID: V2178198
08/29/2023 10:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.9 F 0.0
2:00 176.8 F 2.5
4:00 179.6 F 6.0
6:00 227.1 F 10.8
8:00 220.8 F 4.6
10:00 234.4 F 9.7
12:00 234.2 F 9.3
14:00 241.5 F 11.6
16:00 238.2 F 9.6
18:00 251.3 F 15.7
20:00 262.0 F 22.0
22:00 271.8 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.2
1:00 272.1 F 29.0
2:00 271.9 F 28.5
3:00 272.3 F 29.0
4:00 272.3 F 29.2

Min 271.0 F 28.1
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.8 F 0.0
5:00 194.3 F 0.0
10:00 197.9 F 0.0
15:00 195.6 F 0.0
20:00 191.4 F 0.0
25:00 187.7 F 0.0
30:00 185.7 F 0.0
Drying Complete

FILLING: 1:58
HEATING: 22:27
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:01:51
CYCLE COMPLETE

08/29/2023 11:14 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____