
Midmark M11 - v1.0.0
Total Cycles: 3314

Name: RITTER 1
Sterilizer ID: V2178198
05/31/2023 12:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.7 F 0.0
2:00 177.8 F 0.0
4:00 187.5 F 4.0
6:00 206.0 F 8.7
8:00 222.4 F 13.5
10:00 231.1 F 9.0
12:00 237.8 F 10.2
14:00 243.4 F 12.2
16:00 241.4 F 10.9
18:00 254.9 F 17.8
20:00 266.6 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.5 F 28.7
2:00 271.3 F 28.2
3:00 272.5 F 28.9
4:00 271.6 F 28.2

Min 271.3 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.0 F 0.0
5:00 218.5 F 0.0
10:00 237.5 F 0.0
15:00 235.6 F 0.0
20:00 227.6 F 0.0
25:00 229.6 F 0.0
30:00 230.3 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 21:32
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:59:42
CYCLE COMPLETE

05/31/2023 01:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____