

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2888

Name: RITTER 1  
Sterilizer ID: V2178198  
02/13/2023 08:43 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 75.3 F 0.0  
2:00 79.9 F 0.0  
4:00 112.1 F 0.4  
6:00 128.5 F 2.6  
8:00 149.4 F 5.3  
10:00 171.5 F 8.3  
12:00 193.2 F 12.0  
14:00 206.0 F 3.3  
16:00 216.7 F 6.8  
18:00 227.0 F 10.4  
20:00 232.0 F 8.7  
22:00 237.7 F 10.7  
24:00 234.0 F 7.8  
26:00 247.0 F 14.0  
28:00 257.5 F 19.8  
30:00 267.6 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.6  
1:00 272.6 F 29.3  
2:00 271.5 F 28.6  
3:00 272.6 F 29.4  
4:00 271.7 F 28.7

Min 271.4 F 28.5  
Max 272.6 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.2 F 0.0  
5:00 225.2 F 0.0  
10:00 223.0 F 0.0  
15:00 225.0 F 0.0  
20:00 229.9 F 0.0  
25:00 226.0 F 0.0  
30:00 221.0 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 31:28  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:09:29  
CYCLE COMPLETE

02/13/2023 09:52 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_