
Midmark M11 - v1.0.0
Total Cycles: 1905

Name: RITTER 1
Sterilizer ID: V2178198
05/04/2022 07:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.8 F 0.0
2:00 74.3 F 0.0
4:00 84.3 F 0.0
6:00 118.6 F 0.2
8:00 144.5 F 1.6
10:00 174.1 F 3.9
12:00 194.1 F 7.0
14:00 207.6 F 10.7
16:00 221.6 F 5.9
18:00 226.1 F 7.5
20:00 237.0 F 11.8
22:00 234.4 F 8.3
24:00 243.2 F 11.4
26:00 245.0 F 12.6
28:00 256.9 F 18.9
30:00 267.5 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.6 F 29.0
2:00 271.8 F 28.3
3:00 271.8 F 28.7
4:00 272.5 F 28.8

Min 271.3 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.5 F 0.0
5:00 198.4 F 0.0
10:00 191.6 F 0.0
15:00 199.2 F 0.0
20:00 208.0 F 0.0
25:00 216.0 F 0.0
30:00 225.3 F 0.0
Drying Complete

FILLING: 0:43
HEATING: 31:28
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:09:18
CYCLE COMPLETE

05/04/2022 08:47 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____