
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002585
7:24 01/11/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:23
PRESSURIZING 1:25
STERILIZING 2:25
136.4 C 220kPa 2:25
136.3 C 223kPa 2:55
136.4 C 220kPa 3:25
136.1 C 220kPa 3:55
136.2 C 219kPa 4:25
136.4 C 221kPa 4:55
136.0 C 218kPa 5:25
136.2 C 220kPa 5:55
136.2 C 220kPa 5:55
VENTING 5:55
AIR DRYING 6:46
CYCLE COMPLETE 6:48

Digital Signature #
AC18ADB15FE65AA3
