

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 004168  
15:28 08/08/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:55  
PRESSURIZING 2:58  
STERILIZING 3:56  
136.0 C 220kPa 3:56  
136.4 C 222kPa 4:26  
136.1 C 222kPa 4:56  
136.1 C 221kPa 5:26  
136.2 C 223kPa 5:56  
136.2 C 223kPa 6:26  
136.3 C 224kPa 6:56  
136.4 C 225kPa 7:26  
136.4 C 225kPa 7:26  
VENTING 7:26  
AIR DRYING 8:18  
CYCLE COMPLETE 8:40

-----  
Digital Signature #  
F26BEF24D65094AF  
-----