
Midmark M11 - v1.0.5
Total Cycles: 2835

Name: RITTER2
Sterilizer ID: V2313228
09/06/2023 09:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.8 F 0.0
2:00 179.4 F 2.1
4:00 188.1 F 6.4
6:00 209.3 F 11.8
8:00 226.8 F 7.0
10:00 238.1 F 12.3
12:00 237.7 F 10.1
14:00 236.5 F 8.9
16:00 250.5 F 15.9
18:00 262.8 F 23.0
20:00 272.2 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.4 F 29.1
2:00 272.2 F 29.5
3:00 272.0 F 28.9
4:00 272.5 F 29.6

Min 271.4 F 28.7
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.4 F 0.0
5:00 233.9 F 0.0
10:00 221.2 F 0.0
15:00 210.9 F 0.0
20:00 190.9 F 0.0
25:00 185.2 F 0.0
30:00 208.3 F 0.0
Drying Complete

FILLING: 2:05
HEATING: 20:09
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:59:31
CYCLE COMPLETE

09/06/2023 10:16 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____