

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002413  
12:43 10/01/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:31  
PRESSURIZING 2:37  
STERILIZING 3:46  
136.4 C 226kPa 3:46  
136.4 C 226kPa 4:16  
136.4 C 223kPa 4:46  
136.0 C 222kPa 5:16  
136.3 C 223kPa 5:46  
136.4 C 224kPa 6:16  
136.4 C 226kPa 6:46  
136.3 C 224kPa 7:16  
136.3 C 224kPa 7:16  
VENTING 7:16  
AIR DRYING 8:06  
CYCLE COMPLETE 38:06  
-----

Digital Signature #  
C203784CDD18FAB0  
-----